



# Gloves



# Gloves

At Karex, we continuously endeavour the usage of gloves to prevent any cross-contamination during medical examination. Our gloves are made from the highest quality of raw materials that offers soft texture for improved tactility and better fit.

We provide a range of gloves that are customisable to your requirements.

## WHO IS KAREX?

- ◆ *Karex has the market reputation and proven track record as the world's largest condom manufacturer*
- ◆ *We have the capacity to produce over 5 billion condoms annually*
- ◆ *We export to over 130 countries globally*
- ◆ *We are compliant with various management systems such as ISO 13485, ISO 14001, US FDA 21 CFR Part 820, and GMP*
- ◆ *We also manufacture other products such as personal lubricants, probe covers and foley balloon catheters*

Get in touch with us at [enquiry@karex.com.my](mailto:enquiry@karex.com.my)

Marketing Office :  
Lot 594, Persiaran Raja Lumu, Pandamaran Industrial Estate,  
42000 Port Klang, Selangor, Malaysia  
Tel: +603 3165 6688 Fax: +603 3166 2000

**karex**

# NITRILE GLOVES



3.2 Mil

3.5 Mil

3.5 Mil Extended Cuff

- ◆ Non-sterile
- ◆ 100% synthetic nitrile latex
- ◆ Size: XS, S, M, L, XL
- ◆ Colour: Violet, Skyblue, Black, White

Powder Free

- ◆ Online Chlorinated

## Physical Dimensions

DIMENSIONS	ASTM D6319	EN 455
Length (mm)	min. 220 (S) min. 230 (M,L)	min. 240
Palm Width (mm)		
S	80±10	80±10
M	95±10	95±10
L	110±10	110±10
Thickness Single Wall (mm)		
Finger	min. 0.05	N/A
Palm	min. 0.05	N/A

## Physical Properties

PROPERTIES	ASTM D6319	EN 455
Tensile Strength (MPa)		
Before Aging	min. 14	N/A
After Aging	min. 14	N/A
Elongation At Break (%)		
Before Aging	min. 500	N/A
After Aging	min. 400	N/A
Median Force At Break (N)		
Before Aging	N/A	min. 6
After Aging	N/A	min. 6

- ◆ **SHELF LIFE**  
3 years from the date of manufacturing.
- ◆ **STORAGE**  
Avoid direct sunlight. The gloves shall maintain their properties when stored in a dry location.
- ◆ **QUALITY STANDARDS**  
Conforms to ASTM D6319 and EN 455 standards.
- ◆ **Manufactured under US FDA 21 CFR Part 820 QSR and ISO 13485:2016.**